

6390 Joyce Drive
100
Golden, CO 80403

Phone 303-940-0033
Fax 866-283-0269
www.phenova.com

March 1, 2019

Ms. Kelley Keenan
Environmental Testing Solutions, Inc.
351 Depot Street
Asheville, NC 28801
USA

Dear Kelley,

Thank you for participating in the R23687 Rapid Return™ Proficiency Testing Study. Enclosed is your final report, which has been carefully reviewed by the PT specialists at Phenova. A final report for your laboratory has been sent to all accrediting agencies you requested at the time of your data submittal.

Thank you again for participating in the R23687 Rapid Return™ Proficiency Testing Study. We appreciate working with you and look forward to our next study. If you have any questions, please call us at 866-942-2978.

Report Definitions:

Assigned Value

The Assigned Value is determined from the study mean, gravimetric and volumetric true concentration of an analyte to be analyzed, calculation and/or an appropriate reference value as stipulated in the EPA National Standards for Water Proficiency Testing Studies Criteria Document (current version), the National Environmental Laboratory Accreditation Conference Institute (TNI) criteria (ref: TNI FOT tables, TNI PT Committee) and other documents distributed by accrediting agencies as applicable.

Evaluation Limits

Acceptance Limits are derived from fixed limits, coefficients, constants and calculations stipulated in the EPA National Standards for Water Proficiency Testing Studies Criteria Documents (current version), the National Environmental Laboratory Accreditation Conference Institute (TNI) criteria (ref: TNI FOT tables, TNI PT Committee) and other documents distributed by accrediting agencies as applicable.

Evaluation

<i>Acceptable</i>	The reported value falls within the Acceptance Limits.
<i>Not Acceptable</i>	The reported value falls outside the Acceptance Limits.
<i>No Evaluation</i>	The reported value is non-numeric and can not be evaluated.
<i>NR</i>	As required by the TNI standards and requested by state authorities, any analyte purchased but not reported by your facility is listed as NR (Not Reported).

This report must not be reproduced except in full, without the written approval of Phenova.

Various aspects of Phenova's proficiency testing schemes can from time to time be subcontracted. If subcontracting occurs, it is placed with a competent and qualified subcontractor and Phenova is responsible for all work performed.

This report must not be used to claim product certification, approval, or endorsement by the American Association for Laboratory Accreditation (A2LA) or any agency of the federal government.



6390 Joyce Drive
100
Golden, CO 80403

Phone 303-940-0033
Fax 866-283-0269
www.phenova.com

State Regulatory Agency Request(s):

Phenova has been authorized to send a copy of your R23687 final results to the following state agencies:

NC

Other Regulatory Agency Request(s):

No requests were made to send a copy of your R23687 final results to an agency.

Third Party Request(s):

No requests were made to send a copy of your R23687 final results to a third party.

6390 Joyce Drive Phone 303-940-0033
100 Fax 866-283-0269
Golden, CO 80403 www.phenova.com

Final Report - Rapid Return™

Study: R23687

Opening Date: January 22, 2019 - Closing Date: February 27, 2019

Laboratory: Environmental Testing Solutions, Inc.
351 Depot Street
Asheville, NC 28801
USA

Contact: Ms. Kelley Keenan
828-350-9364

EPA Lab ID: NC01230

Anions 1 (RR-AN1-WS)

Lot #: R23687-08

NELAC Code	Analyte	Analysis Date	Analyst	Method Code	Method Description	Units	Assigned Value	Result	Acceptance Limits	Evaluation
1810	Nitrate as N	2/25/2019	KEK	20113808	SM 4500-NO ₃ ⁻ D-2011	mg/L	5.54	5.54	4.99 - 6.10	Acceptable