

6390 Joyce Drive
100
Golden, CO 80403

Phone 303-940-0033
Fax 866-283-0269
www.phenova.com

March 21, 2013

Ms. Kelley E. Keenan
Environmental Testing Solutions, Inc.
351 Depot Street
Asheville, NC 28801
USA

Dear Kelley,

Thank you for participating in the WPM0213 Water Pollution Proficiency Testing Study. Enclosed is your final report, which has been carefully reviewed by the PT specialists at Phenova. This report presents some additional sections and features which will give you more information on the study data for the standards that were reported by your laboratory and more information regarding the overall performance of your laboratory in relation to other study participants.

The enclosed report is in two sections, the "Evaluation Report" and the "Statistical Report". The "Evaluation Report" contains the evaluation of your reported data as well the reported method, analysis date and analyst. The "Evaluation Report" is an updated and enhanced version of the reports you previously received at the conclusion of our PT studies. The "Statistical Report" contains your passing percentage as well as statistics from the study for the analytes reported by your laboratory.

For any analyte falling outside the established acceptance limits, our PT management staff would like to assist you in determining the most appropriate course of corrective action for your facility. Please contact us at any time if we may be of service to you.

Thank you again for participating in the WPM0213 Water Pollution Proficiency Testing Study. We appreciate working with you and look forward to our next study. If you have any questions, please call us at 866-942-2978.



Report Definitions:

Assigned Value

The Assigned Value is determined from the study mean, gravimetric and volumetric true concentration of an analyte to be analyzed, calculation and/or an appropriate reference value as stipulated in the EPA National Standards for Water Proficiency Testing Studies Criteria Document (current version), the National Environmental Laboratory Accreditation Conference (NELAC) criteria (ref: NELAC FOT tables, NELAC PT Committee) and other documents distributed by accrediting agencies as applicable.

Evaluation Limits

Acceptance Limits are derived from fixed limits, coefficients, constants and calculations stipulated in the EPA National Standards for Water Proficiency Testing Studies Criteria Documents (current version), the National Environmental Laboratory Accreditation Conference (NELAC) criteria (ref: NELAC PT FOT tables, NELAC PT Committee) and other documents distributed by accrediting agencies as applicable.

Evaluation

Acceptable

The reported value falls within the Acceptance Limits.

Not Acceptable

The reported value falls outside the Acceptance Limits.

No Evaluation

The reported value is non-numeric and can not be evaluated.

NR

As required by the NELAC standards and requested by state authorities, any analyte purchased but not reported by your facility is listed as NR (Not Reported).

Study Mean and Standard Deviation

The mean and standard deviation are calculated from the study data using robust statistical calculations when possible. Other statistical calculations may be used if robust statistical calculations are not possible. The displayed values are independent of any statistical calculations required for the calculation of the analyte evaluation limits.

This report must not be reproduced except in full, without the written approval of Phenova.

This report must not be used to claim product certification, approval, or endorsement by the American Association for Laboratory Accreditation (A2LA) or any agency of the federal government.



6390 Joyce Drive
100
Golden, CO 80403

Phone 303-940-0033
Fax 866-283-0269
www.phenova.com

State Regulatory Agency Request(s):

Phenova has been authorized to send a copy of your WPM0213 final results to the following state agencies:

NC

Other Regulatory Agency Request(s):

No requests were made to send a copy of your WPM0213 final results to an agency.

Third Party Request(s):

No requests were made to send a copy of your WPM0213 final results to a third party.

6390 Joyce Drive
100
Golden, CO 80403

Phone 303-940-0033
Fax 866-283-0269
www.phenova.com

Final Report - Water Pollution Proficiency Testing

Study: WPM0213

Opening Date: February 11, 2013 - Closing Date: March 14, 2013

Laboratory: Environmental Testing Solutions, Inc.
351 Depot Street
Asheville, NC 28801
USA

Contact: Ms. Kelley Keenan
828-350-9364

EPA Lab ID: NC01230

Micro (PT-MIC-WP)										Lot #: 5158-12
NELAC Code	Analyte	Analysis Date	Analyst	Method Code	Method Description	Units	Assigned Value	Result	Acceptance Limits	Evaluation
2530	Fecal Coliform, MF	2/12/2013	KEK	20210008	SM 9222 D-1997	CFU/100 mL	774	980	296 - 2020	Acceptable
2500	Total Coliform, MF					CFU/100 mL	652		186 - 2290	NR
2525	E. Coli, MF					CFU/100 mL	755		274 - 2080	NR
2530	Fecal Coliform, MPN	2/13/2013	KEK	20196605	SM 9221 C E-2006	MPN/100 mL	848	1300	228 - 3150	Acceptable
2500	Total Coliform, MPN	2/13/2013	KEK	20191202	SM 9221 B-2006	MPN/100 mL	851	700	231 - 3130	Acceptable
2525	E. Coli, MPN					MPN/100 mL	1060		543 - 2050	NR



6390 Joyce Drive
 # 100
 Golden, CO 80403

Phone 303-940-0033
 Fax 866-283-0269
 www.phenova.com

PT Study Score Report

EPA ID: NC01230
 Laboratory: Environmental Testing Solutions, Inc.
 351 Depot Street
 Asheville, NC 28801
 USA

Study Number	Number of Reported Results	Number of Passing Results	Percent Passing
WPM0213	3	3	100%

Report Definitions:

- Number of Reported Results** The number of results reported which could be evaluated. Results receiving an evaluation of "NR" or "No Evaluation" are not included.
- Number of Passing Results** The number of results reported receiving an evaluation of "Acceptable" or "Check for Error".
- Percent Passing** The percentage of results reported receiving an evaluation of "Acceptable" or "Check for Error".

6390 Joyce Drive
100
Golden, CO 80403

Phone 303-940-0033
Fax 866-283-0269
www.phenova.com

Statistical Summary Report - Water Pollution Proficiency Testing

Opening Date: February 11, 2013 - Closing Date: March 14, 2013

Laboratory: Environmental Testing Solutions, Inc.
351 Depot Street
Asheville, NC 28801
USA

Contact: Ms. Kelley Keenan
828-350-9364

EPA Lab ID: NC01230

Micro (PT-MIC-WP)				Lot #: 5158-12		
NELAC Code	Analyte	Units	Study Mean	Study Standard Deviation	Low Acceptance Limit %	High Acceptance Limit %
2530	Fecal Coliform, MF	CFU/100 mL	795	245	38.2	261
2500	Total Coliform, MF	CFU/100 mL	689	400	28.5	351
2525	E. Coli, MF	CFU/100 mL	420	0.00	36.3	275
2530	Fecal Coliform, MPN	MPN/100 mL	854	423	26.9	371
2500	Total Coliform, MPN	MPN/100 mL	923	417	27.1	368
2525	E. Coli, MPN	MPN/100 mL	1040	264	51.2	193