



6390 Joyce Drive # 100 Golden, CO 80403 Phone 303-940-0033 Fax 866-283-0269 www.phenova.com

March 10, 2021

Ms. Kelley Keenan Environmental Testing Solutions, Inc. 351 Depot Street Asheville, NC 28801 USA

Dear Kelley,

Thank you for participating in the WPM0221 Water Pollution Proficiency Testing Study. Enclosed is your final report, which has been carefully reviewed by the PT specialists at Phenova. This report presents some additional sections and features which will give you more information on the study data for the standards that were reported by your laboratory and more information regarding the overall performance of your laboratory in relation to other study participants.

The enclosed report is in two sections, the "Evaluation Report" and the "Statistical Report". The "Evaluation Report" contains the evaluation of your reported data as well the reported method, analysis date and analyst. The "Evaluation Report" is an updated and enhanced version of the reports you previously received at the conclusion of our PT studies. The "Statistical Report" contains your passing percentage as well as statistics from the study for the analytes reported by your laboratory.

For any analyte falling outside the established acceptance limits, our PT management staff would like to assist you in determining the most appropriate course of corrective action for your facility. Please contact us at any time if we may be of service to you.

Thank you again for participating in the WPM0221 Water Pollution Proficiency Testing Study. We appreciate working with you and look forward to our next study. If you have any questions, please call us at 866-942-2978.

Report Issue Date - 3/10/2021 Report Page - 1 of 6





6390 Joyce Drive # 100

Phone 303-940-0033 Fax 866-283-0269 www.phenova.com

Golden, CO 80403

Report Definitions:

Assigned Value The Assigned Value is determined from the study mean, gravimetric

and volumetric true concentration of an analyte to be analyzed,

calculation and/or an appropriate reference value as stipulated in the EPA National Standards for Water Proficiency Testing Studies Criteria Document (current version), the National Environmental Laboratory Accreditation Conference Institute (TNI) criteria (ref: TNI FOT tables, TNI PT Committee) and other documents distributed by accrediting

agencies as applicable.

Evaluation Limits Acceptance Limits are derived from fixed limits, coefficients, constants

and calculations stipulated in the EPA National Standards for Water Proficiency Testing Studies Criteria Documents (current version),the National Environmental Laboratory Accreditation Conference Institute

(TNI) criteria (ref: TNI FOT tables, TNI PT Committee) and other documents distributed by accrediting agencies as applicable.

Evaluation

Acceptable The reported value falls within the Acceptance Limits.

Not Acceptable The reported value falls outside the Acceptance Limits.

No Evaluation The reported value is non-numeric and can not be evaluated.

NR As required by the TNI standards and requested by state authorities,

any analyte purchased but not reported by your facility is listed as NR

(Not Reported).

Study Mean and Standard Deviation

The mean and standard deviation are calculated from the study data using robust statistical calculations when possible. Other statistical

calculations may be used if robust statistical calculations are not possible. The displayed values are independent of any statistical calculations required for the calculation of the analyte evaluation

limits.





6390 Joyce Drive # 100 Golden, CO 80403 Phone 303-940-0033 Fax 866-283-0269 www.phenova.com

State Regulatory Agency Request(s):

Phenova has been authorized to send a copy of your WPM0221 final results to the following state agencies:

NC

Other Regulatory Agency Request(s):

No requests were made to send a copy of your WPM0221 final results to an agency.

Third Party Request(s):

No requests were made to send a copy of your WPM0221 final results to a third party.

Report Issue Date - 3/10/2021 Report Page - 3 of 6





6390 Joyce Drive

Fax

Phone 303-940-0033 866-283-0269

Golden, CO 80403

100 www.phenova.com

Final Report - Water Pollution Proficiency Testing

Study: WPM0221

Opening Date: February 9, 2021 - Closing Date: March 2, 2021

Laboratory: Environmental Testing Solutions, Inc.

Ms. Kelley Keenan 828-350-9364

Contact:

351 Depot Street Asheville, NC 28801

EPA Lab ID: NC01230 USA

Micro (PT-MIC-WP) Lot #: 5349								: 5349-12		
NELAC Code	Analyte	Analysis Date	Analyst	Method Code	Method Description	Units	Assigned Value	Result	Acceptance Limits	Evaluation
2530	Fecal Coliform, MF	2/11/2021	KEK	20210008	SM 9222 D-2006 (MF)	CFU/100 mL	539	720	173 - 1680	Acceptable
2525	E. Coli, MF					CFU/100 mL	566		209 - 1530	NR
2500	Total Coliform, MF					CFU/100 mL	719		544 - 951	NR
2530	Fecal Coliform, MPN-Multiple Well					MPN/100 mL	736		377 - 1440	NR
2525	E. Coli, MPN-Multiple Well					MPN/100 mL	813		366 - 1810	NR
2500	Total Coliform, MPN-Multiple Well					MPN/100 mL	849		410 - 1760	NR





6390 Joyce Drive

100

Golden, CO 80403

Phone 303-940-0033 Fax 866-283-0269

www.phenova.com

PT Study Score Report

EPA ID: NC01230

Laboratory: Environmental Testing Solutions, Inc.

351 Depot Street Asheville, NC 28801

USA

	Number of	Number of	
Study Number	Reported Results	Passing Results	Percent Passing
WPM0221	1	1	100%

Report Definitions:

Number of Reported Results The number of results reported which could be evaluated. Results receiving an

evaluation of "NR" or "No Evaluation" are not included.

"Check for Error".

Percent Passing The percentage of results reported receiving an evaluation of "Acceptable" or

"Check for Error".





6390 Joyce Drive

100 Golden, CO 80403

Phone 303-940-0033 866-283-0269 Fax www.phenova.com

Statistical Summary Report - Water Pollution Proficiency Testing

Opening Date: February 9, 2021 - Closing Date: March 2, 2021

Laboratory: Environmental Testing Solutions, Inc. Contact: Ms. Kelley Keenan

828-350-9364 351 Depot Street Asheville, NC 28801

EPA Lab ID: NC01230 USA

Micro (PT-MIC-WP) Lot #: 5349-12							
NELAC Code	Analyte	Units	Study Mean	Study Standard Deviation	Low Acceptance Limit %	High Acceptance Limit %	
2530	Fecal Coliform, MF	CFU/100 mL	552	212	32.1	312	
2525	E. Coli, MF	CFU/100 mL	577	151	36.9	270	
2500	Total Coliform, MF	CFU/100 mL	722	66.5	75.7	132	
2530	Fecal Coliform, MPN-Multiple Well	MPN/100 mL	739	170	51.2	196	
2525	E. Coli, MPN-Multiple Well	MPN/100 mL	807	224	45.0	223	
2500	Total Coliform, MPN-Multiple Well	MPN/100 mL	846	199	48.3	207	