



6390 Joyce Drive P # 100 F Golden, CO 80403 w

 Phone
 303-940-0033

 Fax
 866-283-0269

 www.phenova.com

March 12, 2019

Ms. Kelley Keenan Environmental Testing Solutions, Inc. 351 Depot Street Asheville, NC 28801 USA

Dear Kelley,

Thank you for participating in the WPM0219 Water Pollution Proficiency Testing Study. Enclosed is your final report, which has been carefully reviewed by the PT specialists at Phenova. This report presents some additional sections and features which will give you more information on the study data for the standards that were reported by your laboratory and more information regarding the overall performance of your laboratory in relation to other study participants.

The enclosed report is in two sections, the "Evaluation Report" and the "Statistical Report". The "Evaluation Report" contains the evaluation of your reported data as well the reported method, analysis date and analyst. The "Evaluation Report" is an updated and enhanced version of the reports you previously received at the conclusion of our PT studies. The "Statistical Report" contains your passing percentage as well as statistics from the study for the analytes reported by your laboratory.

For any analyte falling outside the established acceptance limits, our PT management staff would like to assist you in determining the most appropriate course of corrective action for your facility. Please contact us at any time if we may be of service to you.

Thank you again for participating in the WPM0219 Water Pollution Proficiency Testing Study. We appreciate working with you and look forward to our next study. If you have any questions, please call us at 866-942-2978.





6390 Joyce Drive Phon # 100 Fax Golden, CO 80403 www.

 Phone
 303-940-0033

 Fax
 866-283-0269

 www.phenova.com

Report Definitions:					
Assigned Value	The Assigned Value is determined from the study mean, gravimetric and volumetric true concentration of an analyte to be analyzed, calculation and/or an appropriate reference value as stipulated in the EPA National Standards for Water Proficiency Testing Studies Criteria Document (current version), the National Environmental Laboratory Accreditation Conference Institute (TNI) criteria (ref: TNI FOT tables, TNI PT Committee) and other documents distributed by accrediting agencies as applicable.				
Evaluation Limits	Acceptance Limits are derived from fixed limits, coefficients, constants and calculations stipulated in the EPA National Standards for Water Proficiency Testing Studies Criteria Documents (current version),the National Environmental Laboratory Accreditation Conference Institute (TNI) criteria (ref: TNI FOT tables, TNI PT Committee) and other documents distributed by accrediting agencies as applicable.				
Evaluation					
Acceptable	The reported value falls within the Acceptance Limits.				
Not Acceptable	The reported value falls outside the Acceptance Limits.				
No Evaluation	The reported value is non-numeric and can not be evaluated.				
NR	As required by the TNI standards and requested by state authorities, any analyte purchased but not reported by your facility is listed as NR (Not Reported).				
Study Mean and Standard Deviation	The mean and standard deviation are calculated from the study data using robust statistical calculations when possible. Other statistical calculations may be used if robust statistical calculations are not possible. The displayed values are independent of any statistical calculations required for the calculation of the analyte evaluation limits.				

This report must not be reproduced except in full, without the written approval of Phenova.

Various aspects of Phenova's proficiency testing schemes can from time to time be subcontracted. If subcontracting occurs, it is placed with a competent and qualified subcontractor and Phenova is responsible for all work performed.

This report must not be used to claim product certification, approval, or endorsement by the American Association for

Laboratory Accreditation (A2LA) or any agency of the federal government.





6390 Joyce Drive # 100 Golden, CO 80403 Phone 303-940-0033 Fax 866-283-0269 www.phenova.com

State Regulatory Agency Request(s):

Phenova has been authorized to send a copy of your WPM0219 final results to the following state agencies: NC

Other Regulatory Agency Request(s):

No requests were made to send a copy of your WPM0219 final results to an agency.

Third Party Request(s):

No requests were made to send a copy of your WPM0219 final results to a third party.





 6390 Joyce Drive
 Phone
 303-940-0033

 # 100
 Fax
 866-283-0269

 Golden, CO 80403
 www.phenova.com

Final Report - Water Pollution Proficiency Testing

Study: WPM0219

Opening Date: February 12, 2019 - Closing Date: March 5, 2019

Laboratory:	oratory: Environmental Testing Solutions, Inc. 351 Depot Street		Ms. Kelley Keenan 828-350-9364	
	Asheville, NC 28801 USA	EPA Lab ID:	NC01230	

Micr	Micro (PT-MIC-WP) Lot #: 5301							: 5301-12		
NELAC Code	Analyte	Analysis Date	Analyst	Method Code	Method Description	Units	Assigned Value	Result	Acceptance Limits	Evaluation
2530	Fecal Coliform, MF	2/13/2019	KEK	20210008	SM 9222 D-2006 (MF)	CFU/100 mL	809	1300	315 - 2080	Acceptable
2525	E. Coli, MF					CFU/100 mL	750		307 - 1840	NR
2500	Total Coliform, MF					CFU/100 mL	878		424 - 1820	NR
2530	Fecal Coliform, MPN	2/13/2019	KEK	20196605	SM 9221 C E-2006 (MPN)	MPN/100 mL	1010	1300	384 - 2640	Acceptable
2525	E. Coli, MPN					MPN/100 mL	1040		556 - 1960	NR
2500	Total Coliform, MPN					MPN/100 mL	1100		386 - 3110	NR





 6390 Joyce Drive
 Phone
 303-940-0033

 # 100
 Fax
 866-283-0269

 Golden, CO 80403
 www.phenova.com

PT Study Score Report

EPA ID:	NC01230
Laboratory:	Environmental Testing Solutions, Inc.
	351 Depot Street
	Asheville, NC 28801
	USA

	Number of	Number of	
Study Number	Reported Results	Passing Results	Percent Passing
WPM0219	2	2	100%

Report Definitions:

Number of Reported Results	The number of results reported which could be evaluated. Results receiving an evaluation of "NR" or "No Evaluation" are not included.
Number of Passing Results	The number of results reported receiving an evaluation of "Acceptable" or "Check for Error".
Percent Passing	The percentage of results reported receiving an evaluation of "Acceptable" or "Check for Error".





6390 Joyce Drive # 100 Golden, CO 80403
 Phone
 303-940-0033

 Fax
 866-283-0269

 www.phenova.com

Statistical Summary Report - Water Pollution Proficiency Testing

Opening Date: February 12, 2019 - Closing Date: March 5, 2019

Laboratory: Environmental Testing Solutions, Inc. 351 Depot Street Asheville, NC 28801

USA

Ms. Kelley Keenan 828-350-9364

EPA Lab ID: NC01230

Contact:

Micr	Micro (PT-MIC-WP) Lot #: 5301-12						
NELAC	Analyte	Units	Study Mean	Study	Low Acceptance	High Acceptance	
Code				Standard	Limit %	Limit %	
				Deviation			
2530	Fecal Coliform, MF	CFU/100 mL	809	253	38.9	257	
2525	E. Coli, MF	CFU/100 mL	673	221	40.9	245	
2500	Total Coliform, MF	CFU/100 mL	901	192	48.3	207	
2530	Fecal Coliform, MPN	MPN/100 mL	993	334	38.0	261	
2525	E. Coli, MPN	MPN/100 mL	1050	218	53.5	188	
2500	Total Coliform, MPN	MPN/100 mL	1110	351	35.1	283	